

## Certificate of Analysis

Product: ISE Diluent  
 Product Number: HT-820638D  
 Shelf Life: 3 Years  
 Storage Conditions: 18 – 25° C  
 Quantity: 6 X 2 L


Parameter	Specified Range	Units
Appearance pH	<b>Clear, free of visible particulate matter</b> 5	pH units

Details
<b>Function:</b> To dilute samples for Na <sup>+</sup> , K <sup>+</sup> , and Cl <sup>-</sup> measurements on the Hitachi 717, 747, 914, and 917 Analyzers.
<b>Quality Specifications:</b> Passed
<b>Reagent Contents:</b> An aqueous solution containing 65 mmol/L Boric acid and preservative.  <b>For <i>in vitro</i> diagnostic use only.</b>

### Traceability

Analyte	Standard	Test Method
pH	<b>Secondary Standard Traceable to NIST</b>	pH Meter

Certified by:



Kathy Fisher  
 Director, Quality Assurance  
 Diamond Diagnostics Inc.