

Certificate of Analysis

Product: ISE Low Serum Standard
Product Number: BK-AUH1014D
Shelf Life: 2 Years
Storage Conditions: 18 – 25° C
Quantity: 4 X 100 mL

Parameter	Specified Range	Units
Appearance	Clear, free of visible particulate matter	
Sodium	130 $\pm$ 1.5	mmol/L
Potassium	3.5 $\pm$ 0.05	mmol/L
Chloride	85 $\pm$ 2	mmol/L

Details

Function: For the quantitative determination of Na⁺, K⁺, and Cl⁻ concentrations in human serum and urine on Olympus Analyzers.

Quality Specifications: Passed


Reagent Contents: An aqueous solution containing 130 mmol/L Na⁺, 3.5 mmol/L K⁺, 85 mmol/L Cl⁻ and preservatives.

For *in vitro* diagnostic use only.

Traceability

Analyte	Standard	Test Method
Sodium	NIST 919b	IL 943 (Flame Photometry)
Potassium	NIST 918b	IL 943 (Flame Photometry)
Chloride	NIST 919b	Corning 925, SAT-500 Salt Analyzer (Titrimetric)

Certified by:



Kathy Fisher
Director, Quality Assurance
Diamond Diagnostics Inc.